

Return & Repair Form:

Customer Information: Boro of Bernardsville

Company Name: Sgt. BRIAN Testa B: _____ S: _____

Date Received: 06-23-11 Date Given to Service: 06-23-11

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-~~7110~~-6510-6810-7510
9510- Drug Tester 5000

Serial #: AR UL - 0046

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR

Whole Instrument

Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD P

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>6808486</u>	<u>Plate</u>	<u>(</u>	<u>n/c w</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes:

FC # ARCA-0908

cal w/ qc ; ops check

Service Technician: _____

Date: 6-28-11