

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Runnemedede Boro P.D. , Runnemedede NJ

Date Received: 05/04/12

Date given to service: 05/04/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR UL-0048

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
6808455	Fuel Cell	1	N/CW
MP CAL 71	Calibration	1	N/C W
6808486	Plate	1	N/CW
MP Labor	Labor	.5	N/C W

Repair Notes: Unit would not connect to the modem , Installed new capacitors at locations C-156 AND C-167.

Replaced defective fuel cell ARYA-0316 with Fuel Cell ARCM-1763.

CAL W/ QC and OPS CHECK

Service Technician JS

Date: 05-08-12