

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER MENDHAM PD Cust # _____
 Date of Return: 5/3 Carrier: FedEx Method: Priority

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0052
 Description: A B PLUS SCREENER _____ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Unit failing LIN-TESTS + SOLD-CHANGE

Part #	Description	Qty	Total
MPCAL71	7110 CAL	1	

Repair Notes: Ran STAD-CHECK and found unit was running high on IR + EC. Calcd unit to repair. Cal off check + Q.C.

Technician: [Signature]

Date: 5/4/04