

RETURN AND REPAIR FORM



Customer Information B: S:
 Company Name: Mendham Twp. P.D., Mendham, N.J.

Date Received: 09/19/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole
 Serial Number: AR UH-0052
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

- 110V A/C Adapter
- Regulator
- Printer Ribbon
- Printer Paper
- Mouthpieces
- 9510 Stylus
- 9510 Top Cover
- Carrying Case
- Dry Gas
- Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp Cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Fuel cell # ARAA-0263 replaced because of leakage, installed fuel cell ARCH-0453.

Cal w/DC; OPS check

Service Technician: [Signature] Date: 9-27-11