

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: Mendham Twp. P.D., Brookside, NJ

Date Received: 02/12/2016

Date given to service: 02/12/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UL-0052

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: April 2016 (54853)

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	5	NC/W
6808455	Fuel Cell	1	NC/W
8315075	Alcotest 7110 Motherboard	1	NC/W
6809500	Cuvette	1	NC/W

Repair Notes: Replaced defective motherboard, cuvette, and fuel cell.

Cuvette was changed to resolve the purging error issue and motherboard was changed to resolve the printing issues.

Fuel Cell = ARHM-0825

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/25/2016