

Dräger safety

Internal Use Only

*please Rush per DOT*

**Return & Repair Form**

**Customer Information:**

CUSTOMER MORRIS County Police Cust # \_\_\_\_\_

Date of Return: 4-26-7 Carrier: FEDEX Method: Standard

**Product Information:**

Product: 7110 7410 6510 6810 Serial # ARUL-0053

Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_

Whole Inst. Top Half Simulator Serial # \_\_\_\_\_

OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

**Accessories:** (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # \_\_\_\_\_
- Printer Paper
- Printer Ribbon
- Casio # \_\_\_\_\_
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

**Repair Information:**

Test # \_\_\_\_\_

Reason for Return: Repair per attached EC Problem

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced leaky F/C. Full cal ops  
check + O.C.

Technician: CRD

Date: 5/1/07