

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER New Jersey Turnpike Authority Cust # _____
 Date of Return: 4-23-7 Carrier: FEDEX Method: Exp. Saver

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0055
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): power cord

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Error 071 see attached

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Couldnt reproduce any of the listed problems. Date and time were corrupt. Set date and time. Verified 60 min count down timer. Full cal ops check + Q.C.

Technician: ARO

Date: 4/24/07