

Return & Repair Form:

Customer Information:

Company Name: NISSP Cranbury NJ B: _____ S: _____
 Date Received: 6 Jun 11 Date Given to Service: 6 Jun 11
 Carrier: FedEx - UPS - DHL - USPS Method: GRD (NDA) - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR Wh-0055
 9510-Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 112

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>n/c w</u>
<u>mp cal 71</u>	<u>cal</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>labor</u>	<u>15</u>	<u>n/c w</u>

Repair Notes:

CLEAR ERROR 112 / MEMORY SITUAS 401.470 / MEMORY LOSS
REPLACE PCB - ✓

cal w/ Qc ; ops check

Service Technician: [Signature]

Date: 6-10-11