

Return & Repair Form

Customer Information:

CUSTOMER Westfield Police Dept Cust # _____
 Date of Return: 1/2/7 Carrier: Fed EX Method: Exp. Saver

Product Information:

Product : 7110 7410 6510 6810 Serial # ARUL-0058
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: repair per attached

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLA30R	LABOR	.5	
MPLA671	7110 CAL	1	

Repair Notes: Replace F/C due to low results on 2nd test during SUMP-CHANGES and subject tests. Full cal ops check + Q.C.

Technician: CPD

Date: 1/5/07