

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER E.G.T. P.D. Cust # _____
 Date of Return: 11/6-4-7 Carrier: FEDEX Method: 2-Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0059
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: High IR Readings

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not verify problem. Full cal
ops check + O.C.

Technician: CRD

Date: 6/8/07