

# Return & Repair Form

## Customer Information:

Company Name: E. Greenwich Twp PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 06-01-2009 Date Given to Service: 06-01-2009  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810 Serial #: AR111 - 0059  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: ERROR 32 + FC?

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808530</u>	<u>SAMPLE SYSTEM</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WARR</u>

## Repair Notes:

REPLACE FC, SAMPLe SYSTEM CAL w/ QC & OPS CHECKS  
(SAMPLe SYSTEM PISTON DAMAGED BY TECH) REPLACED  
AT NO CHARGE

Service Technician: [Signature]

Date: 06-08-09