

# RETURN AND REPAIR FORM



## Customer Information

Company Name: East Greenwich Twp Mickleton NJ

Date Received: 11/01/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  7110  8610  7510  DT5000  7410 Upper-half  9510  7410 Whole

Serial Number: AR ~~AR~~ - 0059  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V AC Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>mp cal 71</u>	<u>Calibrate</u>	<u>1</u>	<u>w/c w</u>
<u>6808955</u>	<u>Fuel Cell</u>	<u>1</u>	<u>w/c w</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C w</u>

Repair Notes: Replaced Fuel cell ARAB-0871, showing signs of leakage, replaced with Fuel cell ARCH10462.

cal w/DC, ops check

Service Technician: [Signature]

Date: 11-2-11