



# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER Svede-Boro PD Cust # \_\_\_\_\_  
 Date of Return: 3/14 Carrier: Fed Ex Method: Priority

### Product Information:

**Product:** 7110 7410 6510 6810 Serial # ARUL-0060  
**Description:** A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
**Whole Inst. Top Half** Simulator Serial # \_\_\_\_\_  
**OTHER:** \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
**Accessories:** (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: CTRL GAS SUPPLY ERROR DURING SENS-CHANGE

Part #	Description	Qty	Total
<u>MRCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not reproduce problem. Cal ops check + QC.

Technician: ASD

Date: 3/16/00