



RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Rowan University P.D., Glassboro Borough, NJDate Received: 01/22/13Date given to service: 01/22/13Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STDProduct: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 WholeSerial Number: AR UL-0060Printer Serial#: ARSim Serial#: DDProbe Serial#: DDWarranty Expires: July 2013Description: A B Plus Demo Screener Trade In

Accessories

 110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel cell	1	NC/W
6808486	Spacer Plates	2	NC/W
6808374	Geared Engine	1	NC/W
MP Labor	Labor	1	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.Fuel Cell ARDK-0625 changed by TMReplaced defective geared engine.CAL W/QC AND OPS CHECKService Technician BC/TMDate: 01/31/13