

# RETURN AND REPAIR FORM



## Customer Information

B: \_\_\_\_\_ S: \_\_\_\_\_

**Company Name:** Rowan University, Glassboro, NJ

**Date Received:** 09/19/14

**Date given to service:** 09/19/14

**Carrier:**  FedEx  UPS  USPS      **Shipping Method:**  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

**Product:**  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

**Serial Number:** AR UL-0060

**Printer Serial#:** AR

**Sim Serial#:** DD

**Probe Serial#:** DD

**Warranty Expires:** \_\_\_\_\_

**Description:**  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 4414161     | Calibration | 1   |            |
| 4414166     | Labor       | .5  |            |
| 6809514     | Breath Hose | 1   |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |

**Repair Notes:** Replaced breath hose due to subject vomiting into the mouthpiece and intake hose of the unit.

**CAL W/QC AND OPS CHECK**

**Service Technician:** BC

**Date:** 09/30/14