

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Cranbury State Police Barracks, N.J.

Date Received: 08/20/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR Wh-0062

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) power cord

## Repair Information

Test#: \_\_\_\_\_

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
8315875	PCB (Board)	1	n/c w
mp Cal 71	Cal	1	n/c w
mp Labor	Labor	.5	n/c w

Repair Notes: Install new PCB (Board), because of two problems found with the circuit board. The arrived with a error (E3), but could not be repaired because of the circuit board issue.

Cal w/ Oc; ops check

Service Technician: [Signature] Date: 7/25/11