

# Return & Repair Form

**Customer Information:**

Company Name: NJ part Authority B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 10-7-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

**Product Information:**

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR UL - 0065  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLATOR</u>		<u>.5H</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE FC / CAL W/ QC & OPS CHECKS -

\_\_\_\_\_


\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Service Technician:



Date:

10-08-08