

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER BELLEVUE Twp PD Cust # \_\_\_\_\_  
 Date of Return: 4/18 Carrier: FedEx Method: Priority

### Product Information:

Product : 7110 7410 6510 6810 Serial # ARUL-0074  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Random Flies on EC

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 Cal</u>	<u>1</u>	<u>n/c</u>
		<u>1</u>	

Repair Notes: Swapped entire EC sample system to eliminate random high/low EC readings. Cal ops check + Q.C.

Technician: CRD

Date: 4/27/06