

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Berlin Twp. P.D., West Berlin, NJ

Date Received: 08/21/14

Date given to service: 08/21/14

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UL-0074

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced fuel cell and spacer plates for preventive maintenance.

Could not duplicate error that was reported. Ran multiple ABA tests and subject tests.

Fuel Cell = ARFD-1126

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 08/25/14