

Return & Repair Form: 10384580
80491751

Customer Information:

Company Name: NEW JERSEY STATE PD B: _____ S: 85003
 Date Received: 1/6/10 Date Given to Service: 1/6/10
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-~~7110~~-6510-6810-7510 Serial #: ARUL-0075
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR _____
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>W/C WARR</u>
<u>MP CAL 41</u>	<u>CAL</u>	<u>1</u>	<u>W/C WARR</u>
<u>MP LABO</u>	<u>LABOR</u>	<u>1.5</u>	<u>W/C WARR</u>

Repair Notes:

INSTALL PCB FROM DEMO UNIT AR MNR-0607
Cal w/Qc & Ops Check

Service Technician: AS

Date: 01/08/2010