

Return & Repair Form:

Customer Information:

Company Name: NJSP B: _____ S: _____
 Date Received: 12-29-08 Date Given to Service: 12-29-08
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

Product Information:

Product: 7410 7110 6510 - 6810 Serial #: AR UL - 0079
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FOEL CELL

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WORK</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WORK</u>
<u>MP LABOR</u>		<u>.5</u>	<u>N/C WORK</u>

Repair Notes:

Replace FC/CAL w/OC's, EAS checks.

Service Technician: [Signature]

Date: 12-30-2008