

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: New Jersey State Police, Newark, NJ

Date Received: 03/11/13 Date given to service: 03/12/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610 Serial Number: AR UL-0079
 7510 DT5000 Printer Serial#: AR
 7410 Upper-half 9510 Sim Serial#: DD
 7410 Whole Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plate	2	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARDM-2227

CAL W/QC AND OPS CHECK

Service Technician BC Date: 03/15/13