

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER VORHEES TWP. POLICE Cust # \_\_\_\_\_  
 Date of Return: 5-2-7 Carrier: FED EX Method: Standard

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0080

Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_

Whole Inst. Top Half Simulator Serial # \_\_\_\_\_

OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter       Regulator       Mag Card Rdr # \_\_\_\_\_
- Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_
- Mouthpieces       Carrying Case       Dry Gas
- Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: See Attached EC Problem

Part #	Description	Qty	Total
<u>6308455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced leaky F/C. Full cal ops check + O2.

Technician: CPO

Date: 5/3/07