

WJ

Return & Repair Form

Customer Information:

CUSTOMER VOORHEES TWP POLICE DEPT. Cust # _____
Date of Return: 8.07.07 Carrier: FX Method: STUD ON.

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0080
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
Probe Serial # _____

OTHER: _____
Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: LOW IR/EC RESULTS

Part #	Description	Qty	Total
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARRANTY</u>

Repair Notes: TESTED 7110 PRIOR TO ANY CAL OR REPAIR - ALL RESULTS WITHIN SPECIFICATIONS - USED 3 DIFFERENT SIMULATOR SETS WITH 2 SOLUTION LOT #'S TO CONFIRM ACCURACY -
CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 08.08.07