

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: Voorhees PD , Voorhees , NJ

Date Received: 07/23/12 Date given to service: 07/23/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UL-0080
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| MP Cal 71 | Calibration | 1 | NC/W |
| 6808455 | Fuel Cell | 1 | NC/W |
| MP Labor | Labor | 1 | NC/W |

Repair Notes: Customer sent solution change report with unit that reported a low IR and EC. With the same fuel cell I ran twenty 0.100 standard check and the results were @ .099 & 0.100.

Installed Fuel Cell number ARDD-1238 for preventive maintenance. Also removed a few spots of cuvette mirrors.

CAL W/QC AND OPS CHECK

Service Technician: JS Date: 07-24-12