

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER Perth Ambay PD Cust # \_\_\_\_\_  
 Date of Return: 3/28/07 Carrier: FedEx Method: Ground

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0081  
 Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: IR + EC problems

Part #	Description	Qty	Total
<u>60808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MP CAL 71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced questionable leaky F/C. Could not reproduce IR drop outs. Full cal ops check + O.C.

Technician: CRD

Date: 3/29/07