

# Return & Repair Form

## Customer Information:

Company Name: New Jersey State PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 6-23-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other \_\_\_\_\_

## Product Information:

Product: 7410 - 110 - 6510 - 6810 Serial #: AR UM-0041  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR \_\_\_\_\_  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	
<u>MPCAL71</u>	<u>CNC</u>	<u>1</u>	
<u>MPLABCR</u>		<u>.5</u>	

Repair Notes:

REPLACED FC / CAL w/ QC & OPS CHECKS

---



---



---



---



---

Service Technician: [Signature]

Date: 06-25-08