

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Bass River Station, New Gretna, NJ

Date Received: 04/22/13 Date given to service: 04/22/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0041
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6809500	Cuvette	1	NC/W
MP Labor	Labor	1	NC/W

Repair Notes: Replaced defective ir sensor in cuvette

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 04/23/13