

# Return & Repair Form:

**Customer Information:** Camden County Back P.D.

Company Name: Camden Back B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 31 May 11 Date Given to Service: 31 May 11  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA (2DY) - 3DY - Other

**Product Information:**

Product: 7410 (7110) - 6510 - 6810 - 7510 Serial #: AR UM - 0094  
 9510 - Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD \_\_\_\_\_ P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>w/c w</u>
<u>6808374</u>	<u>motor</u>	<u>1</u>	<u>n/c w</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>w/c w</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes:  
FC# ARCA-0639  
Replace motor because of squealing noise.  
Cal w/D ; Ops Check

Service Technician: [Signature]

Date: 6-6-11