

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: Camden County P.D., Camden, NJ

Date Received: 03/29/2017 Date given to service: 03/29/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 Serial Number: AR UM-0044
 7110 8610 Printer Serial#: AR
 7510 DT5000 Sim Serial#: DD
 7410 Upper-half 9510 Probe Serial#: DD
 7410 Whole

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC
4414166	Labor	.5	NC
12003	Inlet Valve	1	NC

Repair Notes: Replaced defective Inlet Valve.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/30/17