

Return & Repair Form

Customer Information:

Company Name: BUTLER PD NJ B: _____ S: _____
 Date Received: 6-17-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: ARUM - 0046
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: IR/FC RESULTS

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPMISC</u>	<u>IR FILTER</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCALFI</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE FC / REPLACE IR FILTER (SMALL SPOT ON FILTER)
CAL w/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 06-17-08