

RETURN & REPAIR FORM



CUSTOMER Ephraim Police Dept NJ Just. # _____
 Date ret. 4/19 Carrier Fed Exp Method 2 DAY RMA# _____
 7110 Serial# ARUM 0047 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply): Regulator P.Paper Scotty V Printer Rib. Pouch
 Sim. Temp. Probe Ser.# _____ Keyboard Organizer Key
 Mouthpieces _____ Sim. To cuvette hose Pump to Sim. hose Power cord
 Other (specify) _____
 Other (specify) _____

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
w/ MPCAL 41			1	
		Repair Time		Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:

EC-Sensor was off calibrated causing test failure.
full pc & OPS CHK with calibrated-

Date 21 APR 05 Technician [Signature]

Every above checked item is included Shipping _____