

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Mt. Arlington P.D., Mt. Arlington, NJ

Date Received: 08/30/2016 Date given to service: 08/30/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0049
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	2	
12130	Cermetek	1	

Repair Notes: Replaced defective cermetek chip, c156, and c167.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 10/05/2016