

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Pennsauken Police Cust # _____
 Date of Return: 3/7/07 Carrier: FED EX Method: 2-Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUM-0051
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: repair per attached

| Part # | Description | Qty | Total |
|----------------|------------------|-----------|-------|
| <u>6808455</u> | <u>Fuel Cell</u> | <u>1</u> | |
| <u>MPLABOR</u> | <u>LABOR</u> | <u>.5</u> | |
| <u>MPCAL71</u> | <u>7110 CAL</u> | <u>1</u> | |
| | | | |
| | | | |

Repair Notes: Replaced leaky F/C. Full cal ops
check + Q.C.

Technician: CFD

Date: 3/8/07