

Return & Repair Form:

Customer Information:

Company Name: Princeton Bldg B: _____ S: _____
 Date Received: 7-6 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: ARUM - 0054
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC?

| Part Number | Description | Qty | Total Cost |
|------------------|--------------|----------|-----------------|
| <u>6808455</u> | <u>FC</u> | <u>1</u> | <u>N/A WARR</u> |
| <u>6808486</u> | <u>PLATE</u> | <u>1</u> | <u>N/A WARR</u> |
| <u>MP Cal 71</u> | <u>Cal</u> | <u>1</u> | <u>N/A WARR</u> |
| <u>MP Labor</u> | | <u>3</u> | <u>N/A WARR</u> |

Repair Notes:

REPLACE FC + 1 PLATE
Cal w/PC ; Ops Check

Service Technician: [Signature]

Date: 7-8-09