

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: N. J. S.P., Cranbury, N.J.

Date Received: 01/27/12 Date given to service: 01/27/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR UM-0057  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>0808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp oal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replaced Fuel Cell AR2K-0739, because of leakage  
with Fuel Cell AR2K-0924.

Cal w/DC ops check

Service Technician: [Signature]

Date: 01/31/2012