

5/0 10 249658

Internal Use Only

80326891

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER SOMERDAE Cust # 54983 ⁵⁶⁴⁷⁴

Date of Return: 10 / 31 / 2007 Carrier: UPS DHL USPS Method: GRD NDA 2DY 3DY
OTHER: _____

Product Information:

Product : 7110 7410 6510 6810 Serial # AR UM - 0058

Description: A B PLUS SCREENER _ Printer Serial # AR -

Whole Inst. Top Half Simulator Serial # _____

OTHER : _____ Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

110VAC Adapter Regulator Mag Card Rdr # _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: _____

Part #	Description	Qty	Total
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: REPLACE CELL / CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 11-01-07