

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Berlin Boro P.D., Berlin, NJ

Date Received: 09/02/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR UM-0064  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel cell</u>	<u>1</u>	<u>w/cw</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>w/cw</u>
<u>mp cal 71</u>	<u>Calibrate</u>	<u>1</u>	<u>w/cw</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>w/cw</u>

Repair Notes: Unit arrived with EL system would not recover service report. Installed new Fuel Cell # ARCH-1191.

Cal w/De; ops check

Service Technician: [Signature] Date: 9-8-11