

# Return & Repair Form

## Customer Information:

Company Name: MSR PASS RIVER STATION B S: \_\_\_\_\_  
 Date Received: \_\_\_\_\_ Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR UM - 0072  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: \_\_\_\_\_

| Part Number    | Description     | Qty        | Total Cost      |
|----------------|-----------------|------------|-----------------|
| <u>6808455</u> | <u>FC</u>       | <u>1</u>   | <u>N/C WORK</u> |
| <u>8315075</u> | <u>MAIN PCB</u> | <u>1</u>   | <u>N/C WORK</u> |
| <u>MPCML71</u> | <u>CAL</u>      | <u>1</u>   | <u>N/C WORK</u> |
|                |                 |            |                 |
|                |                 |            |                 |
| <u>MPLABOR</u> | <u>LABOR</u>    | <u>1.0</u> | <u>N/C WORK</u> |

## Repair Notes:

REPLACED FC / MAIN PCB FAILED DURING INITIAL SETUP  
COMPONENT POSITION L-9 VERY HOT - REPLACED PCB - SETUP PCB  
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 6-12-08