

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: N.J.S.P. Bass River Station, New Gretna N.J.

Date Received: 07/29/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0072

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#: _____

Reason for Return: _____

Part Number	Description	Qty	Total Cost
6808374	motor	1	n/c w
6808455	Fuel cell	1	n/c w
mp cal 71	Calibration	1	w/c w
mp labor	labor	.5	n/c w

Repair Notes: Installed a new fuel cell # ARCA-0179

While calibrating unit the motor stop working and the unit showed a error 32, replaced motor.

Cal w/DC ; ops check

Service Technician: [Signature] Date: 8-1-11