

Return & Repair Form

Customer Information:

CUSTOMER MERCHANTVILLE PD Cust # _____
 Date of Return: 3/27 Carrier: Fed Exp Method: 2 DAY

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUM-0073
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Low EC Readings

Part #	Description	Qty	Total
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not see low EC results. Replaced fuel cell as preventative measure. Cal ops check + QC.

Technician: CR

Date: 3/29/00