

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Merchantville Police Dept. Cust # _____

Date of Return: 7/10/06 Carrier: Fed ex Method: Priority

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUM-0073

Description: A B PLUS SCREENER _____ Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Control test failed on BT

Part #	Description	Qty	Total
<u>NPLABOR</u>	<u>LABOR</u>	<u>2.0</u>	<u>N/C</u>
<u>FAXED</u> <u>7/18/06</u>			

Repair Notes: Couldn't reproduce problem. Did SOURCE CHANGE and 25 BT's after manual adjust of IR + EC. with customer sim + probe. Everything looked good. Suspect unit was pulled away from sim to get 6.00 result on ctrl test.

Technician: CRD

Date: 7/18/06