

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Merchantville P.D., Merchantville, NJ

Date Received: 02/10/12 Date given to service: 02/10/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR UM1-0073  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>mp Cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replaced Capacitor at location 6156 to restore modem connection.

Cal w/ QC; ops check

Service Technician: [Signature] Date: 3-9-12