

# Return & Repair Form:

## Customer Information:

Company Name: Somerset Co. Sheriff's Office B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 5-11-10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD (NDA) - 2DY - 3DY - Other

## Product Information:

Product: 7410 (7110) 6510-6810-7510 Serial #: AR WM-0074  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:		Qty	Total Cost
<u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>Labor</u>	<u>1.5</u>	<u>N/C W</u>

Repair Notes:

NEW FET# AR MW-0806

Cal w/ QC / ops check

Service Technician: [Signature]

Date: 05-12-2010