

RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Chesilhurst P.D. , Chesilhurst, NJ

Date Received: 05/01/12 Date given to service: 05/01/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0075
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: Unit would not connect to the modem , Installed a new capacitor at location C-156.

CAL W/ QC and OPS CHECK

Service Technician JS

Date: 05-04-12