

Turn & Repair Form:

10433817-80550503

76551

Customer Information:

Company Name: New Jersey State Police B. OTJA S: _____

Date Received: 9-30-10 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR ILM - 0080
9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR

Whole Instrument Top 1/2 Sim Ser #: _____

Other _____ Probe Ser # DD _____ P _____

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date Jan. 2013

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w/</u>
<u>MP CPL 71</u>	<u>CPL</u>	<u>1</u>	<u>w/c w/</u>
<u>MPLA BOR</u>		<u>.5</u>	<u>N/C w/</u>

Repair Notes:

New FC # ARBE-1129

Cal w/ DC & ops check

Service Technician: [Signature]

Date: 10-1-2010