

Return & Repair Form:

Customer Information:

Company Name: Roseville Park Police B: _____ S: _____
 Date Received: 10/30/09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-~~7110~~6510-6810-7510 Serial #: AR WA-0156
 9510-Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6908486</u>	<u>Plate</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

FC ARUN-0057 Dated 11/2004, Replaced with
FC ARRE-1001

Cal w/ QC & ops check

Service Technician: JS

Date: 11/02/09