

Return & Repair Form:

Customer Information:

Company Name: Rohunoy NS B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3-16-09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx (UPS) - DHL - USPS Method: (GRD) - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 (7110) 6510 - 6810 Serial #: AR WA - 0161  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

Repair Information:

Test #


Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WORK</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WORK</u>
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WORK</u>

Repair Notes:

REPLACED FC -  
CAL W/ QC & OPS CHECKS

Service Technician:



Date: 03/18/2009