RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Rahway P.D., Rahway, NJ

Date Received: 05/29/13 **Date given to service:** 05/29/13

Carrier: ☒ FedEx ☐ UPS ☐ USPS **Shipping Method:** ☐ GRD ☒ 3DAY ☐ 2DAY ☐ NDA-PRI ☐ NDA-STD

Product:
☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WA-0161

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify)

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.
Fuel Cell = AREB-1809

CAL W/QC AND OPS CHECK

Service Technician BC **Date:** 06/04/13

sustxfs010\Public\Forms\return and repair form 07_21_2011.pdf