

# Dräger safety

## Return & Repair Form

**Customer Information:** TWP

CUSTOMER Hamilton P.D. NJ Cust # 70343  
 Date of Return: 1/10/2008 Carrier:  UPS  DHL  USPS Method:  GRD  NDA  2DY  3DY  
 OTHER: \_\_\_\_\_

**Product Information:**

Product 7110 7410 6510 6810 Serial # AR WA - 0162  
 Description: A B PLUS SCREENER Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
**Accessories:** (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

**Repair Information:**

Test # \_\_\_\_\_

Reason for Return: ~~Low FC~~ Low FC RESULTS

Part #	Description	Qty	Total
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCACT1</u>	<u>7110 CAP</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC - PLATE HAD CORROSION - REPLACED

Technician: [Signature]

Date: 01-15-08